



Contribution ID: 16

Type: **Oral presentation**

Updating the BAMline tomography setup

Monday 12 November 2018 18:05 (25 minutes)

The microtomography setup at the BAMline was designed in 2008. The original setup is still in use for CT measurements. With a recent investment into the optics of BAMline, we want to build a pink beam tomography setup by replacing the detector and optics, in order to reduce the scanning time from currently several hours down to minutes. In this talk, the current ideas are presented.

Primary author: Dr GOLLWITZER, Christian (Bundesanstalt für Materialforschung und -prüfung)

Presenter: Dr GOLLWITZER, Christian (Bundesanstalt für Materialforschung und -prüfung)

Session Classification: Session 3