



Contribution ID: 56

Type: **Poster and Speed Talk**

Influence of environmental parameters on calibration drift in superconducting RF cavities

Friday 5 July 2024 10:27 (3 minutes)

Precisely calibrating RF superconducting radio-frequency linear accelerators is crucial for accurately assessing cavity bandwidth and detuning, which provides valuable insights into cavity performance, facilitates optimal accelerator operation, and enables effective fault detection and diagnosis. In practice, however, calibration of RF signals can present several challenges, with calibration drift being a significant issue, especially in settings prone to humidity and temperature fluctuations. In this paper, we delve into the effect of environmental factors on the calibration drift of superconducting RF cavities. Specifically, we examine long-term calibration drifts and explore how environmental variables such as humidity, temperature, and environmental noise affect this phenomenon. The results show that environmental factors, particularly relative humidity, significantly influence calibration drifts. Moreover, we observe and analyze the lag in their influence. By analyzing these correlations, appropriate compensation algorithms can be designed to mitigate and eliminate these effects, thus optimizing calibration accuracy and stability.

Summary

Primary author: SUN, Yue (MSK (Strahlkontrollen))

Co-authors: BELLANDI, Andrea (MSK (Strahlkontrollen)); BRANLARD, Julien (MSK (Strahlkontrollen)); RICHTER, Bozo (MSK (Strahlkontrollen)); SCHMIDT, Christian (MSK (Strahlkontrollen)); EICHLER, Annika (MSK (Strahlkontrollen)); Dr SCHLARB, Holger (DESY)

Presenter: SUN, Yue (MSK (Strahlkontrollen))

Session Classification: Session 3: Controls/Seeding/DAQ

Track Classification: Beam control