



Contribution ID: 238

Type: **Poster DTS**

Soft and hard x-ray spectroscopy with a novel calorimetric superconducting quantum sensor

Monday 3 November 2025 19:00 (3 minutes)

X-ray spectroscopy at synchrotron light sources has emerged as one of the most powerful tools available for the characterization of the chemical, atomic, and electronic properties of materials. Existing x-ray spectrometers provide either excellent energy resolution at low efficiency or moderate energy resolution at high efficiency. Magnetic microcalorimeters (MMCs) may be a “gamechanger” as they promise outstanding energy resolution (already achieving a ΔE_{FWHM} of 1.25 eV at 5.9 keV [1]), a large energy bandwidth, and extremely high detection efficiency. MMCs are ultra-sensitive cryogenic detectors, which rely on converting the energy from incident photons into heat. Using a sensitive thermometer based on the temperature-dependent magnetization of a paramagnetic material in a weak magnetic field, the change in magnetization due to an incoming photon is sensed with a superconducting quantum interference device (SQUID) [2]. We are currently investigating the feasibility of such a x-ray quantum sensor array at multiple synchrotron-radiation beamlines (i.e., X-SPEC, CAT-ACT, INE, and SUL-X) at the KIT Light Source using a universal, compact, and modular platform. The new instrumentation will not only greatly advance the available experimental techniques but also allow for the study of samples containing radionuclide materials with low concentrations, and/or in in situ and operando environments. In this contribution, we will show the current design concept of the detector platform as well as UO_2 measured at the O K-edge with resonant inelastic (soft) x-ray scattering (RIXS) measured at the X-SPEC beamline, which will also be one of the first test materials for the MMC detector.

[1] Krantz, M.; Toschi, F.; Maier, B.; Heine, G.; Enss, C.; Kempf, S. Magnetic Microcalorimeter with Paramagnetic Temperature Sensors and Integrated Dc-SQUID Readout for High-Resolution X-Ray Emission Spectroscopy. *Appl. Phys. Lett.* 2024, 124 (3), 032601. <https://doi.org/10.1063/5.0180903>.

[2] Kempf, S.; Fleischmann, A.; Gastaldo, L.; Enss, C. Physics and Applications of Metallic Magnetic Calorimeters. *J. Low Temp. Phys.* 2018, 193 (3–4), 365–379. <https://doi.org/10.1007/s10909-018-1891-6>.

Speed talk:

Normal speed talk selection

Author: BLANKENSHIP, Mary (Karlsruhe Institute of Technology)

Co-authors: ARLDT, Nik (Institute of Micro- and Nanoelectronic Systems, KIT; Institute for Data Processing and Electronics (IPE), Karlsruhe Institute of Technology (KIT), Germany); WANSORRA, Constantin (Institute for Photon Science and Synchrotron Radiation (IPS), Karlsruhe Institute of Technology (KIT), Germany; Institute for Chemical Technology and Polymer Chemistry (ITCP), Karlsruhe Institute of Technology (KIT), Germany); HAUSCHILD, Dirk (Karlsruhe Institute of Technology (KIT)); STEININGER, Ralph (Institute for Photon Science and Synchrotron Radiation (IPS), Karlsruhe Institute of Technology (KIT), Germany); SANDRELL, Christine (IMS); ADAM, Fabienne; MÜLLER, Michael (Institute of Micro- and Nanoelectronic Systems (IMS), Karlsruhe Institute of Technology (KIT), Germany); VITOVA, Tonya; KEMPF, Sebastian (KIT); WEINHARDT, Lothar (Karlsruhe Institute of Technology (KIT)); HESKE, Clemens (Institute for Photon Science and Synchrotron Radiation, Karlsruhe Institute of Technology)

Presenter: BLANKENSHIP, Mary (Karlsruhe Institute of Technology)

Session Classification: Poster